

OVERVIEW OF HANDHELD REFRACTOMETERS

ORDER NUMBER	HANDHELD REFRACTOMETER
HRB10-T	Handheld refractometer, Brix determination, measurement range 0–10%Brix, measurement accuracy $\pm 0.1\%$ Brix
HRB18-T	Handheld refractometer, Brix determination, measurement range 0–18%Brix, measurement accuracy $\pm 0.1\%$ Brix
HRB32-T	Handheld refractometer, Brix determination, measurement range 0–32%Brix, measurement accuracy $\pm 0.1\%$ Brix
HRB62-T	Handheld refractometer, Brix determination, measurement range 28–62%Brix, measurement accuracy $\pm 0.1\%$ Brix
HRB82-T	Handheld refractometer, Brix determination, measurement range 45–82%Brix, measurement accuracy $\pm 0.1\%$ Brix
HRB92-T	Handheld refractometer, Brix determination, measurement range 58–92%Brix, 38–43 °Baume, 12–27% water content in honey, measurement accuracy $\pm 0.1\%$ Brix, $\pm 0.5\%$ Baume, $\pm 1\%$ water content in honey
HRB90	Handheld refractometer, Brix determination, measurement range 0–95%Brix, measurement accuracy $\pm 0.2\%$ Brix with thermometer 6–36 °C
HRND	Handheld refractometer, measurement of the refractive index, measurement range nD 1.333–1.517, measurement accuracy nD ± 0.0005
HRS10-T	Handheld refractometer, determination of salt content, measurement range 0–10 % salt content, measurement accuracy $\pm 1\%$
HRS28-T	Handheld refractometer, determination of salt content, measurement range 1.3330–1.3834, 0–28 % salt content, measurement accuracy nD ± 0.001 , $\pm 2\%$ salt content
HRM18-T	Handheld refractometer, determination of serum protein, measurement range nD 1.3330–1.3600, 0–12 g/dl, 1.000–1.050 UG, measurement accuracy nD ± 0.0005 , ± 0.2 g/dl, ± 0.002 UG
HRO32-T	Handheld refractometer, measurement of Oechsle, Brix and potential alcohol content, Most scale, measurement range 0–32%Brix, 30–130 °Oe, 4.4–19% alcohol, measurement accuracy $\pm 0.2\%$ Brix, $\pm 1\%$ °Oe, $\pm 0.1\%$ alcohol
HRKFZ-T	Handheld refractometer, battery fluid and anti-freeze tester for ethylene and propylene glycol content, measurement range anti-freeze: -50–0 °C, battery acid: 1.10–1.30 g/cm ³ , measurement accuracy anti-freeze: $\pm 5\%$ °C, battery acid: ± 0.01 g/cm ³

ORDER NUMBER	ACCESSORIES/CONSUMABLES HANDHELD REFRACTOMETER
HRK01	Cover flap for HR series

Version 5.0 | Status: May 2018 | Subject to modifications and errors.