

E061N**CALORIMETER****HEAT OF HYDRATION OF CEMENT**STANDARDS: **EN 196-8** / ASTM C186Comparable to: BS 4550, 1370 / UNE 80102, 7105
DIN 1164 / UNI 7208

Used to determine the heat of hydration of low heat Portland and hydraulic cement.

The apparatus consists of a Dewar flask contained in an insulated material and housed in a wooden box which is hinged so that the flask can be easily removed or replaced.

A “second” hinged wooden box contains the first one, granting a better insulation, as expressly requested by the a.m. Standards.

The Calorimeter is supplied complete with a constant speed electric stirrer, and filler glass funnel.

The standard supply “does not include”:

- the thermometer (to be selected from Beckman or digital model; see accessories)
- the propeller (selecting it from the specific Standard; see accessories)

which must be ordered separately.

Power supply: 230V 1ph 50Hz 150W

Dimensions: 350 x 250 xh 680 mm

Weight : 12 kg approx.

“NEEDED” ACCESSORIES:**E062-02**

BECKMAN centesimal glass mercury thermometer; or:

E062-04

DIGITAL THERMOMETER. Resolution: 0,01 °C.

Complete with probe, or:

E062-04N

DIGITAL THERMOMETER. Resolution: 0,001 °C.

- Memory for 10000 readings
- Displays, stores and prints:
min, max, mean values, delta T
- Alarm if limit values are exceeded
- Battery operated

E061-I1


PROPELLER, conforming to
ASTM C186 Specifications, or:

E061-I2

PROPELLER, conforming to
EN 196-8 Specifications.

ACCESSORY:**V300-19**

Paraffin wax with melting point
55°C to coat the glass parts
which are in contact with the
hydrofluoric acid.
Pack of 4600 g.



E062-02

**SPARE PARTS:**

E062-01 Dewar flask

E062-03 Filler glass funnel

E062-10**Langavant calorimeter**STANDARD: **EN 196-9**

Used to measure the heat of hydration of cements by means of semi-adiabatic method.

The equipment consists of:

Testing calorimeter; calibrated, dia. 160 by 350 mm

Reference calorimeter (same of the testing one), without certificate.
50 mortar box and 20 sand bags.

Measuring system complete with two temperature probes, modem, software to record temperature, analyze and display data with wireless transmission to modem.

To perform the test a PC is required

