

ASTM D 56 – IP 304

SCOPE

Standard Test Method for Flash Point by Tag Closed Cup Tester.

**AUTOMATIC FLASH POINT BY TAG CLOSED TESTER
- MODEL NTA 440 -
- REF 40600 -**

MAIN CHARACTERISTICS

- ✓ Test temperature range: for flash point below 93°C
- ✓ Double Flash detection by ionisation ring and thermocouple
- ✓ Data Storage: 200 results
- ✓ Quick access to calibration parameters
- ✓ Auto diagnostic
- ✓ Liquid bath as per standard method
- ✓ Optional low gas cartridge to work in all types of environments
- ✓ Gas ignition
- ✓ Automatic Gas cut-off at the end of the test
- ✓ Optional cryostat required for samples having a FP below 40°C
- ✓ RS232C output
- ✓ Data acquisition Supervisor Software included possibility to collect the data for a LIMS

SCOPE OF DELIVERY

NTA 440 delivered ready for use with :

- REF 40390** Printer paper for Seiko DPU 414
- REF 40386** Seiko DPU 414
- REF 40614** Pt 100 probe
- REF 40610** Cup without cover
- REF 40347** Cover without probe for NTA 440
- REF 40612** Detection cable
- REF 60214** Insulated tubing for cryostat - 1 m of gas tubing

OPTIONAL ACCESSORIES

Cryostat for test below 40°C

- REF 22216** Cryostat for test < 40°C range -40°C...+200°C : 0,02°C
- REF 40483** Gas cartridge case and 1x 90 ml gas cartridge +
(REF 40484)
- REF 40613** Electric ignitor

CONSUMABLES

- REF 40390** Printer paper for Seiko DPU 414
- REF 40613** Electric ignitor for flame lightening
- REF 14110** Tube for butane propane gas 6 x12 mm (10 m)
- REF 40484** 100 ml Gas cartridge for optional case



ORDERING INFORMATION
REF 40600

NTA 440 for use on AC 230V – 50Hz – 4A

REF 40600/115V

NTA 440 for use on AC 115V - 50/60Hz - 4A

260 x 535x 500 mm (±20 kg)

CONTACT : sales@normalab.com

NORMALAB FRANCE SAS
ZA Caux Multipôles 1 - F-76 190 Valliquerville
Tel. : +33 232 700 100
Fax : +33 232 704 732

www.normalab.com

NTA 440 Leaflet 15

DISTRIBUTED BY

